

## Incision care for proper healing



- Do not swim or soak in a tub for five days.
- Avoid strenuous activities that may pull at the incision or cause a lot of sweating around the insertion area while the incision heals.
- Replace Tegaderm<sup>™</sup> if it becomes saturated; otherwise, leave it on over the Steri–Strips<sup>™</sup>.
- Leave Steri-Strips<sup>™</sup> on until they fall off.
- Trim the edges of the Steri-Strips™ if they start to curl; do not remove them when doing so.

#### Notify your healthcare provider if:

- Your Steri-Strips<sup>™</sup> come off before incision is fully closed.
- You develop a fever or experience pain, redness, swelling, warmth, or drainage at the incision site.

**Note:** If you need to replace your Tegaderm<sup>™</sup>, take extra care so that the Steri-Strips<sup>™</sup> do not get pulled off.

Carefully review the Benefits and Risks section of the Eversense 365 User Guide.

#### **DAY 1: GET STARTED**

- · Read the enclosed User Guide and Quick Reference Guide.
- Download the Eversense 365 App and pair your transmitter to your mobile device.
- Link your sensor and smart transmitter and begin the 24-hour Warm-up Phase.
- Enter the settings provided by your health care provider.
- Turn off your transmitter there is no need to wear it until you begin the Initialization Phase.

## **DAY 2: INITIALIZATION PHASE BEGINS**

- · Place your smart transmitter with adhesive over sensor and complete 4 Initialization calibrations, 2 to 12 hours apart.
- After the 2nd successful calibration, you will begin to see your glucose data.

#### DAY 3 AND BEYOND: DAILY WEAR

- Daily Calibration Phase begins. Calibrate once daily for 13 days.
- · At Day 14 you will calibrate one time per week.
- · To share CGM data with your health care provider, follow the steps in the User Guide

### WHO TO CONTACT WITH QUESTIONS

- · For concerns regarding your insertion site, contact the healthcare provider who did your procedure.
- · For diabetes management questions, contact the healthcare provider who prescribed Eversense.
- For technical questions or help getting started with Eversense, contact 1-844-SENSE4U (736-7348).



# Eversense CGM onboarding videos are available at: www.eversensecgm.com/resources-365

© 2024 Senseonics, Inc. All rights reserved. © 2024 Ascensia Diabetes Care Holdings AG. All rights reserved.



Distributed by Ascensia Diabetes Care 5 Wood Hollow Road Parsippany, NJ 07054 844-SENSE4U (736-7348) eversensecgm.com

Manufactured by Senseonics, Inc. 20451 Seneca Meadows Parkway Germantown, MD 20876-7005 USA The Eversense\* 365 Continuous Glucose Monitoring (CGM) System is indicated for continually measuring glucose levels for up to one year in people (18 years and older) with diabetes. The system is indicated for use to replace fingerstick blood glucose (BG) measurements for diabetes treatment decisions. Fingerstick BG measurements are required for calibration one time a week after day 13, and when symptoms do not match CGM information or when taking medications of the tetracycline class. The sensor insertion and removal procedures are performed by a health care provider. The Eversense 365 CGM System is a prescription device; patients should talk to their health care provider to learn more.

#### For safety information, see bit.ly/eversensesafety

Eversense, Eversense 365 Continuous Glucose Monitoring, and the Eversense logo are trademarks of Senseonics, Incorporated. Other brands and their products are trademarks or registered trademarks of their respective holders and are used solely for informative purposes. No relationship or endorsement should be inferred or implied. PP-SENS-US-0346